



Tibial Shaft

Description	Dimensions			
	<i>L</i>	<i>(W)</i>	<i>(D')</i>	<i>(A)</i>
	cm	mm	mm	mm
550520 Tibial shaft 10 cm	9.9-10.1	10.0-80.0	5.0-50.0	10.0-80.0
550525 Tibial shaft 15 cm	14.9-15.1	10.0-80.0	5.0-50.0	10.0-80.0
550526 Tibial shaft 20 cm	19.9-20.1	10.0-80.0	5.0-50.0	10.0-80.0

Processing Specifications

Wall thickness must be at least 3mm or greater

Measurements

- L overall length
- W Minimum width at mid shaft measured anterior to posterior.
- D' Maximum diameter of the medullary canal measured medial to lateral.
- A Width at mid shaft measured medial to lateral.

